



THE MINIMED™ 670G SYSTEM

SCHOOL NURSE GUIDE

TABLE OF CONTENTS

Purpose & School Orders	2
MiniMed™ 670G System	3
Manual Mode & Auto Mode	3
SmartGuard™ Auto Mode	4
Important information about Auto Mode	4
How to tell when the pump is in Auto Mode	4
Safe Basal	4
A student's responsibilities in Auto Mode	5
Pump Basics	6
Buttons	6
Unlocking and locking the pump	6
Using the pump in Auto Mode	7
With a linked meter — checking BG, calibrating, bolusing	7
With an unlinked meter — checking BG, calibrating, bolusing	9
Recommended Bolus	11
Bolusing for carbs without a BG entry	12
Entering a BG for Auto Mode	13
Temp Target	14
Auto Mode Exits	15
Using the pump in Manual Mode	16
Checking Last Bolus	18
Alerts & Alarms	19
Changing a Battery	21
Things to Remember	22
Appendix: MiniMed™ 670G pump modes	23

PURPOSE & SCHOOL ORDERS

Purpose

This guide is intended to help school nurses with the basic operation of a student's MiniMed™ 670G system. Please note that this booklet does not cover all aspects of insulin pump therapy and continuous glucose monitoring (CGM). It is written for healthcare professionals with some experience with this technology.

School Orders with Backup Plan




Every student on the MiniMed™ 670G system should have signed orders from the student's healthcare professional. They should include:

- Name of device with programmed settings. Note that the pump may be operating in one of two modes. See next page for more information.
- A backup plan if the pump is not able to be used or deliver insulin with:
 - rapid-acting insulin pens or syringes, and doses for food and for correcting high blood glucose (BG)
 - long-acting insulin pens or syringes, and dose if necessary for prolonged stays at school



The responsibilities of the parents/guardians, school nurse, and other school personnel should also be established.

The MiniMed™ 670G System Components



It's a good idea for every student on the MiniMed™ 670G system to have extra pump supplies, i.e., an infusion set, reservoir, and reservoir if the student can change his own infusion set or change with the help of a caregiver or trained professional.

UNDERSTANDING THE MINIMED™ 670G SYSTEM

You may have heard about the MiniMed™ 670G system – the world's first self-adjusting pump system with SmartGuard™ technology.

The MiniMed™ 670G system can be used in two different ways – **Manual Mode** and **Auto Mode**.

Manual Mode is using the pump with or without a continuous glucose monitor (CGM) in a traditional way, like previous pump systems from Medtronic.

Auto Mode, a SmartGuard™ feature, automatically adjusts basal insulin every 5 minutes based on sensor glucose (SG) readings. A student using Auto Mode must still check BGs and calibrate (update) the sensor periodically, as well as bolus for carbs before meals.


HOME SCREEN IN MANUAL MODE



Without CGM

With CGM

HOME SCREEN IN AUTO MODE



SmartGuard™ Auto Mode shields

Manual Mode

Using the pump in a traditional way,

- Basal rates are pre-programmed.
- Bolusing can be done with the Bolus Wizard™ feature or with manual boluses.
- May be used with or without CGM.

Auto Mode

Controlled by a SmartGuard™ algorithm that self-adjusts basal insulin based on sensor glucose readings.

- Basal insulin is automatically adjusted every 5 minutes.
- Bolusing before meals using the Bolus Wizard™ feature is necessary.
- CGM is required.

Important information about Auto Mode:

- Basal insulin is delivered based on SGs.
- Auto Mode uses a target of 120 mg/dL.
- A student can temporarily change the target to 150 mg/dL, like for exercise.
- Carbs must be entered into the pump before meals.
- BG checks are necessary to calibrate the sensor.
- When a student enters a BG over 150 mg/dL, Auto Mode may recommend a correction bolus.
- A student may receive a BG required alert if the pump needs a BG for Auto Mode.

How to tell when the Pump is in Auto Mode



SmartGuard™ Auto Mode shield


If you see the SmartGuard™ shield, the pump is in Auto Mode, which includes Safe Basal.

Safe Basal

There are times in Auto Mode when basal insulin is being delivered according to *recent* insulin needs, but is not being adjusted based on SG readings. This is called **Safe Basal**, and when the pump is in Safe Basal, you will see the screen to the right.

Safe Basal activates, for example, if the pump and transmitter are not communicating, or a BG entered is very different from the SG. Very often, these situations will resolve themselves before the student is aware of it.

If the pump is in Safe Basal and there is something that can be done to resolve the issue, the pump will alert the student what to do, like check a BG.



A Student's Responsibilities in Auto Mode

When a student is wearing the MiniMed™ 670G system and Auto Mode is active, the student must still perform certain tasks:

1. Check BG and calibrate sensor
2. Bolus for carbs before eating
3. Respond to alarms and alerts

Check BG and calibrate sensor

Students should check their BGs prior to meals and calibrate their sensor. Calibrating the sensor is performing a fingerstick, and using that BG value to update the device. It's best to calibrate the sensor 3-4 times a day, like before meals and bedtime. So while at school, it's reasonable a student might calibrate once before lunch and/or if the device asks for a calibration.

Bolus for carbs before eating

When in Auto Mode, a student must bolus for carbs before each meal and snack. Giving insulin before a meal can help students avoid post-meal highs, which could lead to fewer alerts and improved glucose control. Check the school orders for the ideal length of time to bolus pre-meal.

Respond to alarms and alerts

Students should respond promptly to all alarms and alerts to help spend more time in Auto Mode and avoid highs and lows, which could lead to more time spent in target range.

USING AUTO MODE



Sam's pump is in Auto Mode, and she would like to eat a meal. She knows that she should check her BG, calibrate the sensor and then enter carbs into the pump for the food she is certain she will eat.

Note: indicated for ages 7 and over.

Pump Buttons



Backlight

When you are not pressing buttons on the pump, you will notice that the Backlight will soon turn off. The pump is still on; it is just saving battery life. You can simply press any button to make the screen reappear.

Unlocking the Pump

After the Backlight has been off for a few minutes, the pump goes into Sleep mode and the pump is locked.

To use the pump, press Select twice. You will see a screen like the one shown here. Press the arrow key that is highlighted to unlock the pump.



Locking the Pump

If you would like to lock the pump, simply press and hold the Graph button.

USING THE PUMP IN AUTO MODE

Using the CONTOUR®NEXT LINK 2.4 meter to enter a BG with or without carbs for food, calibrate the sensor, and deliver a bolus*



In Auto Mode (with a linked meter)
1. Check BG
2. Calibrate sensor
3. Enter carbs and bolus

1. Check BG.
2. Select **Yes** to confirm the BG meter reading.

If you do not believe the meter result is accurate, do not confirm now. Select **No**, wash hands, and recheck BG.



3. **Bolus** will be highlighted. If you want to calibrate with this BG, select **Calibrate Sensor**.



4. Select **Yes** to calibrate.
Select **No** to not calibrate.



5. If you want to give a bolus, select **Bolus**.
If you do not want to give a bolus, press **✓** and select **Done**.




*Do not calibrate your CGM device or calculate a bolus using a blood glucose meter result taken from an Alternative Site (palm) or from a control solution test. It is not recommended to calibrate your CGM device when sensor or blood glucose values are changing rapidly, e.g., following a meal or physical exercise.

USING THE PUMP IN AUTO MODE

6. Select **Carbs** to enter carbs for food.

If you are not eating carbs, go to the next step.



7. Select **Next** to review the calculated bolus amount.



8. Select **Deliver Bolus** to give the bolus.



The Bolus Started message briefly appears, then the Home screen appears, with a banner showing the bolus being delivered.






USING THE PUMP IN AUTO MODE

To manually enter a BG and carbs for food, deliver a bolus, and calibrate the sensor:

1. Press .
2. Select **Bolus**.
3. Select **BG**.
4. Press  or  to enter your BG reading, and press .
5. Select **Carbs**.

In Auto Mode (with an unlinked meter)

1. Check & enter BG
2. Enter carbs & bolus
3. Calibrate sensor



USING THE PUMP IN AUTO MODE


6. Press **^** or **v** to enter carbs for your food, and press **O**.
7. Select **Next**.



8. Review the calculated bolus amount.
9. Select **Deliver Bolus** to deliver the bolus.




The message Bolus Started briefly appears.



A message appears asking if you want to calibrate using the entered BG.

10. Select **Yes** to calibrate.
Select **No** to not calibrate.



The Home screen appears showing the bolus being delivered.




USING THE PUMP IN AUTO MODE

Recommended Bolus

If a BG entered is greater than 150 mg/dL, Auto Mode may recommend a correction bolus.

1. Read the message on the first screen.
2. Press  to finish reading the message.



3. Select **Bolus**.

Auto Mode will calculate how much insulin to deliver.



A CORRECTION BOLUS IN AUTO MODE




Sam's pump is in Auto Mode. After lunch, Sam notices her glucose levels are higher than normal, so Sam checks her BG. Auto Mode recommends a correction bolus for the high BG value, and Sam delivers the bolus. Sam feels safe knowing that Auto Mode estimates her correction bolus amount based on her glucose needs at that moment.

USING THE PUMP IN AUTO MODE

Bolusing for carbs without a BG entry

There may be times in Auto Mode when a student would like to eat a second helping of food or a snack without checking a BG.

1. Press .
2. Select **Bolus**.
3. Press  to **Carbs** and press .
4. Press  to enter the amount of carbs you are eating and press .
5. Select **Next**.
6. Select **Deliver Bolus**.



The Home screen appears showing the bolus being delivered.




USING THE PUMP IN AUTO MODE

Entering a BG

There may be times that Auto Mode requests a BG entry. You may check with a linked meter, or manually enter the BG.

If you are manually entering the BG:

1. Press  from the Home screen.
2. Select **Enter BG** and manually enter BG value.
3. Select **Save**.



TEMP TARGET

To Enter a Temp Target

The standard Auto Mode target is 120 mg/dL, although a student may want to temporarily change the Auto Mode target to 150mg/dL, like for physical activity.

1. Press .
2. Select **Temp Target**.
3. Press  or  to set the Temp Target duration and then press . The duration can be set in 30 minute increments. The default is 2 hours.
4. Select **Start**.




The message Temp Target Started briefly appears, then the Home screen appears with a banner showing the remaining Temp Target time.



To Cancel a Temp Target

To return to the standard Auto Mode target of 120 mg/dL before the Temp Target duration expires, a student can cancel the Temp Target.

1. Press .
2. Select **Cancel Temp Target**.
The Temp Target screen appears and shows the details of the temp target.
3. Select **Cancel Temp Target** to cancel the temp target.





The Temp Target Ended message and duration of the Temp Target briefly appear. Then the Home screen appears.

Auto Mode Exits

Why do Auto Mode exits occur?

There are times when the pump will exit Auto Mode and return to Manual Mode for safety reasons.

There could be an alarm that needs attention. For example, a student could have a high SG (over 300 mg/dL) for more than 1 hour. The pump has exited to Manual Mode, and will need a BG to return to Auto Mode. This exit to Manual Mode allows a student time to troubleshoot the hyperglycemia and take action to resolve it, like give a correction bolus and monitor glucose levels.



What to do if there is an Auto Mode exit?


The Auto Mode Readiness screen (shown at right) helps you or a student determine why Auto Mode is not active. Go to the screen to see if there is something the student can do to activate Auto Mode, such as check a BG. The pump may ask you if you would like to view the Auto Mode Readiness screen. Or you can go there from the Home screen by selecting:

[Menu > Status > Auto Mode Readiness](#)



Manual Mode

If there is an Auto Mode exit, the pump will go into Manual Mode. In Manual Mode, a student's pre-programmed Basal rates will start automatically. Also, a student can use the Bolus Wizard™ feature to bolus for meals and corrections.



When in doubt, call the Medtronic 24-hour helpline for device-related questions, or the student's parent or caretaker.

USING THE PUMP IN MANUAL MODE

How to use the Bolus Wizard™ feature for Manual Mode

Deliver correction and food bolus

1. Check BG.
2. Press **○**.
3. Select **Bolus**.
4. Select **Bolus Wizard**.

If using a linked meter, the **BG** is on screen. If not, select **BG**.

5. Press **^** or **v** to enter BG and press **○**.
6. Select **Carbs**.
7. Press **^** to enter grams of carbs and press **○**.
8. Select **Next**.
9. Select **Deliver Bolus**.



Deliver correction bolus—no food

1. Check BG.
2. Press **○**.
3. Select **Bolus**.
4. Select **Bolus Wizard**.

If using a linked meter, the **BG** is on screen. If not, select **BG**.

5. Press **^** or **v** to enter BG and press **○**.
6. Press **v** and select **Next**.
7. Select **Deliver Bolus**.



USING THE PUMP IN MANUAL MODE

Deliver food bolus—no correction

1. Press **○**.
2. Select **Bolus**.
3. Select **Bolus Wizard**.
4. Press **✓** and select **Carbs**.
5. Press **^** to enter the amount of carbs you are eating and press **○**.
6. Select **Next**.
7. Select **Deliver Bolus**.



For more information about using the MiniMed™ 670G in Manual Mode, go to: www.medtronicdiabetes.com and click on **Support**

MANUAL MODE



Sam is taking steroids short-term for an illness, so Sam's doctor instructed Sam to use her pump in Manual Mode until she is done with the steroids and her insulin needs have returned to her typical daily doses. Sam feels good that she can still benefit from her insulin pump features and continuous glucose monitor.

CHECKING LAST BOLUS

Checking Last Bolus

Whether a student's pump is in Auto Mode or Manual Mode, there may be times when you need to see the time or amount of the last bolus that was given. For example, you may want to check to make sure a student took a bolus at lunch. You can see the last bolus delivered in the **Quick Status** screen.

1. Press .
2. Press  to **Status** and press .
3. Press  to **Quick Status** and press .



Quick Status		9:00 AM
Last bolus	2.800 U (N)	9:02 AM
		Jan 1
Last BG	265 mg/dL	9:03 AM
		Jan 1

Checking Bolus History

You may also want to review the last several boluses that were delivered. For example, a parent might want to know the boluses their child gave throughout the day. You can see the last several boluses delivered in **Daily History**.

1. Press .
2. Press  to **Options** and press .
3. Press  to **History** and press .
4. Press  to **Daily History** and press .
5. Press  on the day you would like to review.






Daily History		9:00 AM
Bolus (N)	0.050 U	3:32 PM
Bolus (N)	0.100 U	2:07 PM
Bolus (N)	0.050 U	11:55 AM
◀ Fri, Mar 30 ▶		

ALARMS & ALERTS




Here are some common alarms & alerts that you might see on a student's pump in Auto Mode and/or Manual Mode, and how to respond.

Read and address the alert, then clear it by pressing  then 

Alert	Reason	Steps to take
	A new BG entry is required for Auto Mode.	Perform fingerstick and enter a new BG.
	Auto Mode recommends a correction bolus based on a BG that you have entered.	Consider delivering the recommended correction bolus.
	A calibration is required to keep your pump in Auto Mode.	Perform a fingerstick. Enter BG and calibrate your sensor.
 <p><i>Followed by</i></p> 	SG has been high for over one hour. This value is based on a set glucose threshold and length of time: 300 mg/dL or higher for one hour; 250 mg/dL or higher for three hours.	<p>High SG</p> <p>Check infusion set. Check ketones. Monitor BG.</p> <p>Auto Mode Exit</p> <p>Monitor BG and treat as necessary. Enter BG to continue in Auto Mode.</p>
	SG is under 50 mg/dL.	Perform fingerstick and treat as needed. Monitor BG.


ALARMS & ALERTS

Read and address the alert, then clear it by pressing  then 


Alert	Reason	Steps to take
	The sensor is updating	Do not calibrate unless notified. This could take up to 3 hours.
	Your system was unable to use the BG you entered to calibrate your sensor	In 15 minutes, your pump will prompt you to enter a new BG for calibration. Wash hands before checking.
	Low battery	Change battery when possible. See next page for how-to instructions.
	Failed battery test	Try again, or change battery and use new battery.
	Low Reservoir	Change reservoir when possible.
	Insulin Flow Blocked	Read message on screen to understand the alarm and choose the desired option.

CHANGING THE BATTERY


The pump is powered by a AA battery. A brand new Lithium, Alkaline, or fully-charged rechargeable battery can be used.



1. Unscrew the battery cap using the bottom edge of the belt clip. (Or use a thick coin.)




2. Insert battery with negative (flat) end going in first.



3. Place battery cap into the pump and use the edge of the belt clip to screw the cap back on.

Do not under-tighten or try to over-tighten the battery cap. It should be aligned horizontally with the pump case as shown here.



Battery Alerts

- Low battery pump alert — 8-10 hours of battery life remains
- Replace battery alert — 30 minutes of battery life remains
- Replace battery now alarm — insulin delivery stopped due to low power

THINGS TO REMEMBER

The MiniMed™ 670G system with SmartGuard™ technology can help keep your students' glucose levels in target range.* More time spent in target range may help your student live a healthier life and focus on learning!

Things to remember in Auto Mode:

- ✓ A student must check BGs and calibrate the sensor, bolus before meals, and respond to alarms and alerts
- ✓ Highs and lows can still occur, so make sure to have a plan in place on how to address them



For any urgent technical questions, please call the **Medtronic 24-hour helpline** at **1-800-646-4633, option 1**.





For additional information & support, go to www.medtronicdiabetes.com

Other Helpful Resources:

American Diabetes Association — www.diabetes.org

JDRF (Juvenile Diabetes Research Foundation) — www.jdrf.org

*Bergenstal R, Garg S, Weinzimer S, et al. Safety of a hybrid closed loop insulin delivery system in patients with type 1 diabetes. *JAMA*. 2016; 316(13): 1407-1408.

	Manual Mode	Auto Mode	Auto Mode-Safe Basal
Home screen display	  <p>Without CGM With CGM</p>		
Availability	When Auto Mode is not active	Auto Mode is active	Pump automatically transitions to Safe Basal when SGs are available
Basal	Uses the basal settings programmed into the pump	Automatically adjusts basal insulin every 5 minutes depending on the SG value	A fixed rate is delivered for a maximum of 90 minutes. If the cause doesn't resolve, then pump exits to Manual Mode
Bolus	Uses the programmed Bolus Wizard™ settings to recommend a dose	Must enter carbs into pump. Uses carb ratio to recommend a bolus. Auto Mode calculates and recommends a correction if BG >150 entered	Same as Auto Mode - Bolus

IMPORTANT SAFETY INFORMATION

WARNING: (For MiniMed™ 670G System Users Ages 7-13): The low sensor glucose alert functionality is distinct from the automated insulin dosing function of the MiniMed™ 670G system. When used in Auto Mode, the MiniMed™ 670G system has been shown to be safe and effective for its intended use in this population. However, do not rely solely on the use of a low sensor glucose (SG) value for "Alert on Low" or "Alert before Low" for alerts set at 50 mg/dL and 60 mg/dL. A low sensor glucose alert may not reflect the user's true blood glucose at these levels or may not alert. Do not ignore symptoms of low glucose. Always confirm your sensor glucose readings with your blood glucose meter and treat according to the recommendations of your healthcare professional. Solely relying on these sensor glucose alerts and readings for treatment decisions could result in missing severe hypoglycemia (low blood glucose) events.

IMPORTANT SAFETY INFORMATION MINIMED™ 670G SYSTEM

The Medtronic MiniMed™ 670G system is intended for continuous delivery of basal insulin (at user selectable rates) and administration of insulin boluses (in user selectable amounts) for the management of type 1 diabetes mellitus in persons, seven years of age and older, requiring insulin as well as for the continuous monitoring and trending of glucose levels in the fluid under the skin. The MiniMed™ 670G system includes SmartGuard technology, which can be programmed to automatically adjust delivery of basal insulin based on Continuous Glucose Monitor sensor glucose values, and can suspend delivery of insulin when the sensor glucose value falls below or is predicted to fall below predefined threshold values. The system requires a prescription. The Guardian™ Sensor 3 glucose values are not intended to be used directly for making therapy adjustments, but rather to provide an indication of when a fingerstick may be required. A confirmatory finger stick test via the CONTOUR®NEXT LINK 2.4 blood glucose meter is required prior to making adjustments to diabetes therapy. All therapy adjustments should be based on measurements obtained using the CONTOUR®NEXT LINK 2.4 blood glucose meter and not on values provided by the Guardian™ Sensor 3. Always check the pump display to ensure the glucose result shown agrees with the glucose results shown on the CONTOUR®NEXT LINK 2.4 blood glucose meter. Do not calibrate your CGM device or calculate a bolus using a blood glucose meter result taken from an Alternative Site (palm) or from a control solution test. It is not recommended to calibrate your CGM device when sensor or blood glucose values are changing rapidly, e.g., following a meal or physical exercise. If a control solution test is out of range, please note that the result may be transmitted to your pump when in the "Always" send mode.

WARNING: Medtronic performed an evaluation of the MiniMed™ 670G system and determined that it may not be safe for use in children under the age of 7 because of the way that the system is designed and the daily insulin requirements. Therefore this device should not be used in anyone under the age of 7 years old. This device should also not be used in patients who require less than a total daily insulin dose of 8 units per day because the device requires a minimum of 8 units per day to operate safely.

Pump therapy is not recommended for people whose vision or hearing does not allow recognition of pump signals and alarms. Pump therapy is not recommended for people who are unwilling or unable to maintain contact with their healthcare professional. The safety of the MiniMed™ 670G system has not been studied in pregnant women. For complete details of the system, including product and important safety information such as indications, contraindications, warnings and precautions associated with system and its components, please consult <http://www.medtronicdiabetes.com/important-safety-information#minimed-670g> and the appropriate user guide at <http://www.medtronicdiabetes.com/download-library>

Medtronic

18000 Devonshire Street
Northridge, CA 91325
USA
1.800.646.4633
(24-hour technical support for
physicians and medical professionals)

www.medtronicdiabetes.com

950M17343-01.2 NOV2018 © 2018 Medtronic. All rights reserved. Medtronic, Medtronic logo and Further, Together are trademarks of Medtronic. Third party brands are trademarks of their respective owners. All other brands are trademarks of a Medtronic company. PP-CNL-2-4-US-0564. Ascensia, the Ascensia Diabetes Care logo, and Contour are trademarks and/or registered trademarks of Ascensia Diabetes Care.